

## InNovaSound & Integrity USB Ultrasound Systems

### Minimum computer requirements

- Operating Systems - Windows XP, Vista or Mac\*
- Minimum processor - 1 GHz
- Minimum RAM - 512 MB RAM
- One (1) USB 2.0 port
- Minimum Display\* 1280x800 resolution with 32bit color \*
- Laptop displays must include XBRITE technology and NVidia® Ge Force™ Graphics System

\*Macs require Windows emulator "BootCamp" and either Microsoft XP or Vista

For any questions or clarifications call Direct Medical Systems Technical Support at

888 600 8666 - option 1 - in the USA

925 600 8463 - option 1 - outside of the USA